



TELEDYNE
ODOM HYDROGRAPHIC

DIGI-BAR Test and Calibration certificate

Valid for surface use*

Type : DIGIBAR PRO Date of issue : 03-11-2020
Serial No. 003780 Variant: PRO

: sound velocity meter

Calibration Constants	
VK1	-185
VK2	10164
VK3	31
PK1	2
PK2	6167
PK3	31

Calibration & Final Function Test :

Sign :

* Surface use: 0 to 50m water depth.



TELEDYNE
ODOM HYDROGRAPHIC

TELEDYNE-RESON A/S, Fabriksvangon 13, DK-3660
Slangerup Fax: +45 4738 0066, Phone: +45 4738 0022

**STANDARD
DEL
GROSSO
H2O**

**SERIAL#
98XXX**

TEMP	VELOCITY	MEASURED FREQUENCY	RES VEL	OBS-CAL	TEMP	VELOCITY	MEASURED FREQUENCY	RES VEL	OBS-CAL
4.0	1421.62		1421.4		17.5	1474.38		1474.6	
4.5	1423.90		1423.7		18.0	1476.01		1476.2	
5.0	1426.15		1426.0		18.5	1477.62		1477.9	
5.5	1428.38		1428.3		19.0	1479.21		1479.4	
6.0	1430.58		1430.5		19.5	1480.77		1481.0	
6.5	1432.75		1432.7		20.0	1482.32		1482.6	
7.0	1434.90		1434.9		20.5	1483.84		1484.1	
7.5	1437.02		1437.0		21.0	1485.35		1485.6	
8.0	1439.12		1439.1		21.5	1486.83		1487.0	
8.5	1441.19		1441.2		22.0	1488.29		1488.5	
9.0	1443.23		1443.2		22.5	1489.74		1490.0	
9.5	1445.25		1445.3		23.0	1491.16		1491.4	
10.0	1447.25		1447.3		23.5	1492.56		1492.7	
10.5	1449.22		1449.3		24.0	1493.95		1494.1	
11.0	1451.17		1451.3		24.5	1495.32		1495.4	
11.5	1453.09		1453.2		25.0	1496.66		1496.8	
12.0	1454.99		1455.1		25.5	1497.99		1498.1	
12.5	1456.87		1457.0		26.0	1499.30		1499.3	
13.0	1458.72		1458.9		26.5	1500.59		1500.6	
13.5	1460.55		1460.7		27.0	1501.86		1501.8	
14.0	1462.36		1462.6		27.5	1503.11		1503.1	
14.5	1464.14		1464.3		28.0	1504.35		1504.2	
15.0	1465.91		1466.1		28.5	1505.56		1505.4	
15.5	1467.65		1467.9		29.0	1506.76		1506.6	
16.0	1469.36		1469.6		29.5	1507.94		1507.8	
16.5	1471.06		1471.3		30.0	1509.10		1508.9	
17.0	1472.73		1472.9						