



**TELEDYNE**  
**ODOM HYDROGRAPHIC**

## DIGI-BAR Test and Calibration certificate

Valid for surface use\*

Type :	DIGIBAR PRO	Date of issue :	07-10-2019
Serial No.	003780	Variant:	PRO

Calibration Constants	
VK1	-164
VK2	10141
VK3	31
PK1	0
PK2	6155
PK3	31

Calibration & Final Function Test :

Sign :

QA Signat ure:

Sign :

\* Surface use: 0 to 50m water depth.



**TELEDYNE**  
**ODOM HYDROGRAPHIC**

TELEDYNE-RESON A/S, Fabriksvangen 13, DK-3550  
Slangerup Fax: +45 4738 0066, Phone: +45 4738 0022

STANDARD  
DEL  
GROSSO  
H2O

SERIAL#  
98XXX

TEMP	VELOCITY	MEASURED FREQUENCY	RES VEL	OBS-CAL	TEMP	VELOCITY	MEASURED FREQUENCY	RES VEL	OBS-CAL
4.0	1421.62		1421.4		17.5	1474.38		1474.6	
4.5	1423.90		1423.7		18.0	1476.01		1476.2	
5.0	1426.15		1426.0		18.5	1477.62		1477.9	
5.5	1428.38		1428.3		19.0	1479.21		1479.4	
6.0	1430.58		1430.5		19.5	1480.77		1481.0	
6.5	1432.75		1432.7		20.0	1482.32		1482.6	
7.0	1434.90		1434.9		20.5	1483.84		1484.1	
7.5	1437.02		1437.0		21.0	1485.35		1485.6	
8.0	1439.12		1439.1		21.5	1486.83		1487.0	
8.5	1441.19		1441.2		22.0	1488.29		1488.5	
9.0	1443.23		1443.2		22.5	1489.74		1490.0	
9.5	1445.25		1445.3		23.0	1491.16		1491.4	
10.0	1447.25		1447.3		23.5	1492.56		1492.7	
10.5	1449.22		1449.3		24.0	1493.95		1494.1	
11.0	1451.17		1451.3		24.5	1495.32		1495.4	
11.5	1453.09		1453.2		25.0	1496.66		1496.8	
12.0	1454.99		1455.1		25.5	1497.99		1498.1	
12.5	1456.87		1457.0		26.0	1499.30		1499.3	
13.0	1458.72		1458.9		26.5	1500.59		1500.6	
13.5	1460.55		1460.7		27.0	1501.86		1501.8	
14.0	1462.36		1462.6		27.5	1503.11		1503.1	
14.5	1464.14		1464.3		28.0	1504.35		1504.2	
15.0	1465.91		1466.1		28.5	1505.56		1505.4	
15.5	1467.65		1467.9		29.0	1506.76		1506.6	
16.0	1469.36		1469.6		29.5	1507.94		1507.8	
16.5	1471.06		1471.3		30.0	1509.10		1508.9	
17.0	1472.73		1472.9						